



Technical Data Sheet—Date: 10/25/2011 (Page 1 of 2)

Intersorb 408 Granules

Product: Intersorb 4 to 8 mesh Indicating and Non-indicating soda lime Grade D

Description:

Product name: Intersorb 408 NI (non-indicating) Intersorb 408 WV (indicating)

Properties:

Intersorb 408 is comprised of 2 mm cylindrical granules and has been produced to achieve the maximum carbon dioxide absorption and

optimum physical properties. This is to achieve the most suitable performance within rebreathers.

Intersorb 408 has been tested to NATO test standard STANAG No 1411.

Application / Use:

Rebreather, Recreation Rebreather Diving, Commercial Diving, Military, Mine Safety Equipment.

Specifications:

Chemical Composition: Intersurgical tests.

	Intersorb 408 NI	Intersorb 408 WV
Calcium Hydroxide	97%	97%
Sodium Hydroxide	3%	3%
Ethyl Violet	NIL	0.03%

Note: These figures represent the dry constituents. The product will additionally contain 14% to 18% water.

Physical properties: NATO test standard STANAG No 1411

Thysical properties. NATO test standard OTAMAO NO 1411				
Particle Size	Intersorb 408	Specification		
	NI and WV			
	Typical data			
Over 2.80 mm	0.6%	1% max		
2.00 to 2.80 mm	25%	30% max		
1.40 to 2.00 mm	Balance	Balance		
0.600 to 1.40 mm	6%	20% max		
Under 0.600 mm	0.5 %	1% max		
Moisture content	16 %	14 % to 20 %		





Technical Data Sheet—Date: 07/11/2013 (Page 2 of 2)

Intersorb 4 to 8 mesh Indicating and Non-indicating soda lime Grade D **Product:**

Carbon Dioxide absorption: NATO test standard STANAG No 1411

Particle Size	Intersorb 408	Specification
I di tiole dize	NI and WV	Opecification
	Typical data	
Hardness	87%	80% minimum
(% Retained on 1.4mm		
screen)		
Resistance to flow	1.4 mbar unused	
(40 L/min, absorber 10 cm	1.6 mbar used	
diameter, 12.5 cm height,		
volume 1 litre.)		
Time to 0.5 %	100 minutes	80 minimum
CO ₂		
breakthrough		
(minutes)		
CO ₂ capacity L/kg	150 L/kg	120 L/kg minimum

105 ml absorbent in 30 mm diameter tube.

Challenge gas: 3.0 L/min air containing 5 % CO₂.

Humidity 100 %